

## **DECLARATION**

Labbox Labware declares under the manufacturer responsibility that this good:

HMAD-100; HMAD-1K0; HMAD-250; HMAD-2K0; HMAD-500

is in conformity with the:

Directive 2014/30/EU - EMC (European Electromagnetic Compatibility)

Directive 2014/35/EU - LVD (Low Voltage Directive)

Directive 2011/65/EU amended by 2015/863/EU RoHS (Restriction of Hazardous Substances)

Moreover, this equipment has tested against the standards:

EN 61326-1: Electrical equipment for measurement, control and laboratory use – EMC requirements – Part 1: General Requirements.

EN 61000-3-2: Electromagnetic Compatibility (EMC) Part 3-2: Limits for harmoniccurrent emissions (Equipment input current up to and including 16A per phase)

EN 61000-3-3: Electromagnetic Compatibility (EMC) Part 3-3: Limitation of voltage supply systems for equipment with rated current ≤ 16A

EN IEC 62321-2:2021, EN 62321-3-1:2014, EN 62321-5:2014, EN 62321-4:2014/A1:2017, EN 62321-7-1:2015, EN 62321-7-2:2017, EN 62321-6:2015, EN 62321-8:2017

This document has been produced electronically and it is valid without signature: 05/09/2024